



# Certificate of Traceability

## Level I Laboratory Certification


This certifies that ICON Central Laboratories, using Bio-Rad Variant II has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#7

The system evaluated was:

Instrument: <b>Variant II</b>	Calibrator Lot: AN92576NU, AN92577NU	Column Lot: NU62000P
Reagent Lot: AA00193NU, AA00194NU, AA00195NU	Calibrator Assigned Values: 5.6%, 10.3%	

Date of Certification: August 1, 2010

Certification Expires: August 1, 2011



NGSP Steering Committee Chair



NGSP Network Coordinator



SRL director/ supervisor